



Stretchable Composite Acoustic Transducer for Wearable Monitoring of Vital Signs

Yasin Cotur,* Michael Kasimatis, Matti Kaisti, Selin Olenik, Charis Georgiou, and Firat Güder*

A highly flexible, stretchable, and mechanically robust low-cost soft composite consisting of silicone polymers and water (or hydrogels) is reported. When combined with conventional acoustic transducers, the materials reported enable high performance real-time monitoring of heart and respiratory patterns over layers of clothing (or furry skin of animals) without the need for direct contact with the skin. The approach enables an entirely new method of fabrication that involves encapsulation of water and hydrogels with silicones and exploits the ability of sound waves to travel through the body. The system proposed outperforms commercial, metal-based stethoscopes for the auscultation of the heart when worn over clothing and is less susceptible to motion artefacts. The system both with human and furry animal subjects (i.e., dogs), primarily focusing on monitoring the heart, is tested; however, initial results on monitoring breathing are also presented. This work is especially important because it is the first demonstration of a stretchable sensor that is suitable for use with furry animals and does not require shaving of the animal for data acquisition.

1. Introduction

In modern global healthcare, both for humans and animals, wearable devices are expected to play a major role for remote and continuous monitoring of health and the early detection of diseases.^[1] The increasing cost of healthcare, limited access to health centers (e.g., due to long waiting times), and aging populations are the main drivers of wearable health monitor development.^[2] The recent advances in wearable monitoring tools, however, were only made possible with the emergence of low-cost mobile devices (e.g., smartphones).^[3] When combined with mobile devices, wearable health monitors can acquire, manipulate, store, and transmit data inexpensively compared to conventional

solutions.^[4] At the same time, flexible and stretchable devices could provide unobtrusive and continuous monitoring of health without sacrificing comfort.^[5–12]

(Opto)electronic sensors integrated into flexible/stretchable materials,^[13,14] with a wearable form factor, offer an accurate yet inexpensive method to monitor vital signs of health (heart rate in particular) noninvasively.^[15] The biggest shortcoming concerning these sensors (such as wearable optoelectronic heart rate monitors), however, is that they must be worn over the bare-skin of the user for reliable collection of data. Users may even need to shave parts of their body to be able to make accurate measurements.^[16] This produces inconvenience, discomfort, and often causes users to stop wearing their devices (since they continuously need to shave). Photoplethysmography (PPG) and electrocardiography (ECG) are two popular methods used

in wearables for measuring heart rate (HR). While PPG is an optical method that detects HR by way of detecting the rate of blood flow, ECG captures electrical signals generated by the contraction and relaxation of the heart muscles.^[17,18] Both methods generally require hair-free skin; often, compensation for motion artefacts is also needed. Most wearables for measuring heart rate require direct contact with the skin. Materials with low compliance do not conform to the curvature of the body; therefore, this produces poor-quality or faulty data, the devices are less comfortable and may require additional materials (such as conductive gels) for the measurement.^[19] There are also remote methods of measuring HR,^[20] such as ballistocardiography and seismocardiography, and breathing,^[21] e.g., radio-frequency respiration monitoring, which provide unobtrusive collection of data but they fail to function when subjects move, due to motion artefacts.

The use of stretchable materials with high compliance for the fabrication of wearable devices provides the means for creating robust yet comfortable tools for continuous monitoring of vital signs such as heart rate.^[22] Conventional electronics can be encapsulated with soft materials to improve conformation to the curvature of the skin and may be worn like a tattoo.^[23] Conformal and direct contact with the skin using wearable tattoos also enables alternative modes of monitoring heart rate including acoustic or other pressure-based sensing modalities.^[24,25] These devices, however, still require direct (hard) contact with the skin and generally fail when the tissue is wet;

Y. Cotur, M. Kasimatis, Dr. M. Kaisti, S. Olenik, C. Georgiou, Dr. F. Güder
Department of Bioengineering
Imperial College London
London SW7 2AZ, UK
E-mail: y.cotur16@imperial.ac.uk; guder@imperial.ac.uk

 The ORCID identification number(s) for the author(s) of this article can be found under <https://doi.org/10.1002/adfm.201910288>.

© 2020 The Authors. Published by WILEY-VCH Verlag GmbH & Co. KGaA, Weinheim. This is an open access article under the terms of the Creative Commons Attribution License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

DOI: 10.1002/adfm.201910288

they may also cause irritation/allergic reactions. Furthermore, sophisticated “tattoo-like” stretchable devices require advanced methods of microfabrication (hence expensive) and are meant to be thrown away after a single use (i.e., disposable).

In this article, we report a stretchable wearable composite acoustic transducer for monitoring vital signs. The composite device uses water or hydrogels as the medium for the propagation of acoustic waves which are encapsulated inside a stretchable silicone membrane (note that water/hydrogels are not applied directly to the skin). The acoustic signals originating from within the body are captured using battery-operated conventional microelectronics that can monitor HR and respiration over long periods, store, analyze and transmit the data to a nearby computer wirelessly. Unlike most wearable health monitors, our system can be worn over clothing or hairy skin (coat/fur) of animals such as working, military or police dogs without the need for shaving the animal. In this work, we particularly focus on monitoring HR (and only provide preliminary results for monitoring respiration) and characterize the system developed for both dog and human subjects.

2. Results and Discussion

2.1. Fabrication of the Transducer

Figure 1 illustrates the four-step fabrication procedure for the water–silicone composite acoustic transducer. In this scheme, i) a polylactic acid (PLA)-based polymer mold is 3D printed to cast the bottom-outer layer of the silicone membrane (2 mm in thickness). ii) The silicone membrane produced is filled with deionized water (or hydrogels) and, iii) uncured silicone is poured directly on top of the water to fully encapsulate it (see Video S1 of the Supporting Information for the production of

water–silicone composite transducer). During this process, the silicone introduced naturally spreads itself over the water added inside the bottom-outer silicone membrane (i.e., does not mix) and forms another thinner membrane (20–100 μm) which covers the entire top-surface and is highly stretchable and flexible (Figure S1, Supporting Information). This process, though it may appear counterintuitive, works extremely effectively and allows the silicone membrane to cure and bond with the top edges of the bottom layer. This method also works if the deionized water is replaced with a hydrogel (Figure S2, Supporting Information), producing a hydrogel–silicone composite. The use of hydrogels may reduce or fully eliminate the evaporation of water, however at the expense of increased weight and slight reduced flexibility. Once cured, the resulting structure contains no air bubbles (air bubbles reduce the quality of data acquired acoustically). iv) In the final step, a conventional microphone with its associated electronics is placed on the outside of the silicone membrane and encapsulated with an additional layer of silicone to produce a monolithic device (Figure S3, Supporting Information). To create a wearable form-factor, the water–silicone transducer can also be monolithically integrated into a silicone-based stretchable harness (Figure 1) which contains additional custom-built, battery-operated wireless electronics (Figure S3, Supporting Information). The final system produced can be worn over the chest and conforms to the curvature of the body, allowing acquisition of acoustic signals from the subject.

2.2. Characterization of Performance

We investigated the performance of the composite acoustic transducers for monitoring heart sounds, i.e., phonocardiography (PCG). PCG is a powerful method that produces a graphical recording of all sounds and murmurs originating from the




Figure 1. A) Fabrication steps of the water–silicone composite transducer: (1) Degassed, uncured liquid silicone is poured in the mold and left to cure partially for 2 h. (2) Partially cured silicone is removed from the mold and filled with water. (3) Uncured, liquid silicone is poured; silicone spreads itself over the water and continues to cure with the partially cured part, fully encapsulating the water. (4 and 5) Microphone is placed in the recess and buried in more silicone to create a monolithically wearable harness. B) Photograph of the wireless electronics, battery, and microphone. C) Photograph of the harness produced by embedding the microphone amplifier with silicone. Electronics and battery are placed in a 3D printer container and placed in the sleeve on the harness.

flow of blood and motion of valves in the heart.^[26] There are two major heart sounds: i) the lub sound—S1—arises from the closure of mitral and tricuspid valves in the beginning of systole; ii) the dub sound—S2—occurs during the closure of aortic and pulmonic valves.^[27] A successful PCG device should be able to resolve these sounds accurately with little noise (most PCG devices are susceptible to movement), which can be used for measuring HR or making diagnosis of diseases of the heart.^[28–30]

In the first experiment, we characterized the composite transducers using simulated heart sounds (see Figure S4, Supporting Information, for the experimental setup). To benchmark the quality of the sounds recorded with composite transducers and understand the effects of materials/geometry on transduction, we produced seven different types of silicone composite transducers (see Figure 2A for time series data) using the following materials: i) commercial, nonflexible, and nonstretchable stethoscope diaphragm, ii) air–silicone composite with 15 mm height—this material is similar to the water–silicone composite but instead is not filled with water and contains air, iii) all-silicone transducer with 15 mm height—consists of a block of silicone, iv) water–silicone composite with 15 mm and v) 30 mm height of water. We downloaded a reference PCG recording of a healthy heart beating at 60 beats min^{-1} from an online repository by ThinkLabs Medical LLC.^[31] Most heart sounds are observed between 20 and 100 Hz as shown in the frequency response of the original sound signal (see Figure S5, Supporting Information, for original heart sound). We played this sound using a loudspeaker and rerecorded the sounds generated using the transducers (see Figures S6 to S9 of the Supporting Information for time series and Figures S10–S13 of the Supporting Information for the frequency spectra of the recordings).

Visual evaluation of the time series data alone indicates that the air–silicone composite and all-silicone transducers produce recordings with lower quality; the air–silicone transducer attenuates the signal of interest and S2 is barely visible due to the noise. The all-silicone transducer creates little attenuation but introduces significant noise to the recording, reducing overall quality. To produce a quantitative measure of similarity, we used dynamic time warping (DTW) to calculate the error between the reference recordings that is played back, and the signal captured by the transducers. DTW (see Description S1 of the Supporting Information for more information on DTW) is widely used as a similarity measure especially for biological signals such as speech.^[32–34]

The DTW analysis shown in Figure 2B indicates that the water–silicone composite transducers ($n = 7$) produce the smallest mean error among all stretchable materials, in comparison to the reference recording. Although the all-silicone transducers have the lowest variability, they exhibited the worst overall performance. Water–silicone composite transducers with a height of 30 mm and air–silicone composite transducers produced signals with higher mean error in comparison to the 15 mm water–silicone composite transducer. The transducer with the commercial stethoscope diaphragm performed the best among all devices tested; this result was expected since the attenuation of acoustic waves is lowest in dense solids. Of course, the commercial device is not stretchable, does not




Figure 2. A) Recording of simulated heart sounds. The simulated heart sounds (reference recording downloaded from an online repository) were played through loudspeakers and rerecorded using transducers made of different materials. Air–silicone and all-silicone transducers were both 15 mm in height whereas for water–silicone composite transducers, the heights were 15 and 30 mm; B) Error between the rerecorded sounds, shown in (A), and reference recording, calculated using dynamic time warping (DTW). C) Normalized and filtered waveforms from (A) – Pass-band: 20–100 Hz. D) Error between the bandpass filtered, rerecorded normalized sounds, shown in (C), and bandpass filtered, normalized reference recording calculated using DTW. E) Error between rerecorded sounds and original recording calculated using DTW as a function of strain.

conform to the human body, and is susceptible to motion artefacts (moves easily), thus it is not suitable for continuous wearable sensing.

The degree of attenuation that an acoustic wave experiences depends on two factors: i) the acoustic impedance which is

the product of the density of the medium of propagation and speed of sound in that medium,^[35] ii) the coupling between the mediums of propagation if the waves are traveling across an interface.^[36] Unstable contact between (solid) mediums also produces noise that reduces the quality of a signal. Although in terms of the acoustic impedance of the medium, the all-silicone transducer had an advantage over the other silicone-based devices because it is less flexible and less stretchable compared to the air–silicone and water–silicone composite transducers, the coupling between the media was poorer due to the reduced contact area. This resulted in significant error during transduction (see Video S2 of the Supporting Information for comparison of stretchability). The 15 mm water–silicone composite transducer performed the best, most possibly due to the following three reasons: i) higher flexibility/stretchability due to the thin silicone membrane compared to all-silicone transducers, providing better coupling between the transducer and loudspeaker, ii) water provides a medium of propagation with a lower acoustic impedance compared to air,^[37] and iii) increasing the height of the water–silicone composite transducer from 15 to 30 mm also increases the length of propagation the acoustic waves need to travel from the surface of the speaker to the microphone, reducing signal-to-noise-ratio. Since the silicone–water composites have a positive Poisson's ratio, we did not reduce the thickness lower than 15 mm as this would bring the microphone too close to the surface when stretched which could induce noise.

2.3. Enhancement of Transduction Performance Using Digital Filters

We applied a bandpass filter (infinite impulse response Chebyshev filter with a passband of 20–100 Hz; Figure 2C) to the as recorded signals. Through visual inspection of the time series data, it is obvious that after filtering, the signal acquired from the all-silicone transducer remained noisy. The amplitude of the signals recorded with the air–silicone composite transducer exhibited an increase in quality and the S1 and S2 could now be clearly identified. The quality of the recordings measured by the 15 mm water–silicone composite transducer resembled closely the original heart sound and the 30 mm composite had higher noise. Once again, using DTW we have quantitatively measured the degree of similarity between the original recording and the measurements made with the transducers produced (Figure 2D). Digital filtering improved the quality of all recordings: the 15 mm performed the best, similar to the commercial stethoscope diaphragm but with the added benefit of being stretchable/flexible, which would provide improved coupling when worn over clothing or a patch of hairy skin. In the remaining experiments, therefore, we used the water–silicone composite transducer with 15 mm thickness.

2.4. Effect of Strain

To understand the effect of stretching on the performance of the 15 mm water–silicone composite transducer, we placed the devices to be tested on a loudspeaker (Harman Kardon Onyx

Studio 4) with a curved surface. We stretched seven samples of water–silicone composite transducers at six levels of strain up to 75 mm to measure error with DTW (Figure 2E). The DTW results show that stretching does not produce a significant change in the performance of the water–silicone composite transducers, although there is a slight increase in variability/error when the devices are stretched without a trend. Nonetheless, we can conclude that stretching has a negligible effect on performance.

2.5. Testing with Healthy Human Volunteers and Dogs

In the next experiment, we tested the 15 mm water–silicone composite transducer to record the heart sounds of healthy human volunteers ($n = 5$) wearing their daily clothing. For example, the subject shown in Figure 3A was wearing three layers of clothing on the day of the experiment (see Figure S14 of the Supporting Information for comparison of acquired signals versus layers of clothing). All recordings (see Video S3, Supporting Information) were made in a standard laboratory environment with ambient noise. The heart sounds recorded before and after digital filtering, while the subject was standing, are shown in Figure 3B. We included the waveforms recorded by a commercial stethoscope for comparison. Since the commercial stethoscope was rigid and did not conform to the body well, the amplitude of the signals recorded were low, and the heart sounds were difficult to detect. Even though we were able to recover the S1 sounds by bandpass filtering (Butterworth filter with a passband of 20–150 Hz), the S2 sounds were still not visible. By contrast, both the S1 and S2 sounds of the heart could easily be identified with the water–silicone composite transducer. We have also made recordings using the conventional stethoscope and water–silicone composite transducer over the bare skin of a human volunteer (Figure S15, Supporting Information). These measurements produced comparable signals between both devices if the volunteer stayed still however movement introduced substantial noise similar to the measurements over clothing. With the stretchable composite transducer, the recordings had a much better signal-to-noise profile (Figure S16, Supporting Information) most probably due to the improved contact (i.e., improved conformation to the shape of the body of the subject). We did not observe a significant change in the quality of the signal recorded with the water–silicone composite transducer before and after filtering. We also applied wavelet denoising and envelope detection algorithms to obtain heart rates from the waveforms recorded during sitting or standing (Figure 3C; Figure S17, Supporting Information). For further validation of our method we measured the electrical activity of the heart using conventional ECG (note that the ECG electrodes were attached directly on the skin) while making PCG recordings with the water–silicone composite transducer. As shown in Figure 3D, both recordings produced highly correlated signals as reflected by the heart rates determined during the measurements (60 bpm).

In addition to that of human volunteers, we also tested the water–silicone composite transducer for monitoring the heart




Figure 3. A) Testing with healthy human volunteers. The water–silicone composite transducer is worn over layers of clothing. B) As recorded sounds from the body with the water–silicone composite transducer versus commercial stethoscope—left. After bandpass filtering with a Butterworth filter (passband 20–150 Hz), the sound quality improves further—right. C) Algorithmic detection of S₁ and S₂ waveforms recorded from a human subject and subsequent identification of heart rate. D) Simultaneous recording of ECG (using commercial electrodes attached directly on the skin) and PCG (recorded with the water–silicone composite transducer) signals showing functional agreement.

rate of a dog (**Figure 4A**). The test subject was a healthy Labrador Retriever. Unlike humans, most dogs have a thick coat of fur, rendering our method of measuring heart sounds particularly suitable for monitoring the heart rate of furry animals. Figure 4B shows the sounds of the heart of the dog which were bandpass filtered (passband: 20–150 Hz) to remove unwanted signals—, e.g., breathing sounds, ambient noise. Both the S₁ and S₂ regions can be clearly identified in the resulting waveform.

Overall, the wearable water–silicone composite transducers performed well with healthy human and animal subjects and allowed recordings of PCG continuously without the need for shaving or using conductive gels.




Figure 4. A) Testing of water–silicone composite transducers worn over the furry coat of a dog. B) Algorithmic detection of S₁ and S₂ waveforms recorded from the dog and subsequent identification of heart rate.

3. Conclusions

The water–silicone composite acoustic transducer is an unobtrusive, wearable measurement tool that can be used for monitoring vital signs continuously. Although in this work we have mainly focused on measuring heart sounds, as the spectral distribution of frequencies belonging to breathing and heart sounds/murmurs are different from each other, the breathing patterns can also be measured through the use of digital filters (Figure S18, Supporting Information).^[38]

The system is sufficiently low-cost (\$0.5 for silicone, \$30 for the electronics, battery, and microphone) compared to a commercial electronic stethoscope and easy to fabricate using widely accessible prototyping tools (e.g., 3D printers). The use of an Internet-of-Everything system based on the inexpensive ESP32 board allows acquisition and transmission of recordings to a remote device wirelessly over long periods of time (see Figure S19 of the Supporting Information for a recording made over 24 h). The digital recordings can be preprocessed using the on-board microcontroller and further processed on a nearby (more powerful) mobile device with cloud storage capabilities. Machine learning algorithms can be applied to the data collected to detect anomalies and alert healthcare professionals/veterinarians (or pet owners) of illnesses.

The wearable system, in its current format has at least four disadvantages: i) silicone elastomers require curing (i.e., slow) and are not compatible with low-cost mass manufacturing methods such as injection molding. ii) Acoustic transduction is susceptible to external environmental sounds which reduce the quality of the recordings although the contribution of the external sounds could be reduced through further sound-proofing. iii) When used with animals such as dogs the system has difficulty collecting high quality heart sounds from dogs that pant strongly. iv) Sudden movement of animals (especially highly active dogs) may also introduce additional noise to the recordings through physical contact with external objects or

body parts (such as legs). These disadvantages can, however, be largely eliminated by the use of additional sensors to measure breathing/panting (such as elastomeric strain sensors recently developed in our group)^[39] and a different form-factor for the harness.

The system proposed is suitable for use with humans and expected to work with a range of animals such as furry pets, working animals (such as sniffer dogs) or livestock although in this work, we only performed testing with a dog. In the future, the wearable composite device presented may provide an unobtrusive alternative to current health monitors for human and animal care at a lower cost.

4. Experimental Section

Noninvasive Testing with Human Volunteers and Dog: Prior to noninvasive human and dog experiments, a detailed risk assessment was performed according to the guidelines provided by Imperial College London. All human volunteers and dog were healthy. No signs of stress were observed from the dog during the experiments as the subject was already trained to wear a harness.

Fabrication of Transducers and Harness: Silicone elastomer Ecoflex 30 (produced by Smooth-On, USA) was used in the fabrication of the transducers by mixing Part A and Part B in a 1:1 weight ratio. Once mixed, the mixture was degassed for 15 min in a vacuum chamber to remove any air bubbles before pouring into a mold and curing at room temperature.

Water-Silicone Composite Diaphragm: The bottom part of the transducer was produced using a 3D printed PLA mold. The silicone piece was released from the mold after 2 h of curing and filled with deionized water. Note that after 2 h, the Ecoflex 30 could only partially cure as intended. 5 g of uncured Ecoflex 30 mixture was poured on the surface of the water filled in the bottom silicone piece. The Ecoflex 30 spreads itself over water on its own, reaching the edges of the bottom silicone piece, and continues curing for another 4 h into a composite structure. It is noticed that the silicone-water composite transducer loses small amounts of water over time (hence slightly decreased in size), potentially due to the porosity of silicone materials. Although this did not alter the experimental results over several months, this may be rectified by using liquids/gels with lower vapor pressure.

Air-silicone Composite Diaphragm: The bottom section of the air-silicone diaphragm was created by using the same procedure as described above. The top part of the diaphragm was created using a 3D printed mold that produces a silicone layer of 2 mm in thickness. Uncured Ecoflex 30 was poured into this new mold and a part-cured bottom section was placed in contact. The curing continued for another 4 h.

All-Silicone Diaphragm: Fabrication of the all-silicone piece was straight forward and involved replacing of water in the case of water-silicone composite, with uncured Ecoflex 30 that was poured into the part-cured Ecoflex 30 bottom piece. The entire piece was cured for another 4 h to produce the final structure.

Assembly of Transducers and Harness: The final transducer structure was produced by placing a small printed circuit board (PCB) containing an electret microphone and a MAX9814 amplification chip on the bottom of the diaphragm and encasing the PCB once again with Ecoflex 30. Ecoflex 30 was also used to produce a harness with a 3D printed mold which was again cocured with the transducer to create a monolithically integrated wearable device.

Mold Production: The PLA molds used in the experiments were designed in SolidWorks and printed using a Raise3D N2 Plus 3D printer. The PLA filaments were purchased online from 3dgbire. The geometry

of the molds was modeled after the Littman Classic II S.E. commercial stethoscope with a diameter of 45 mm.

Electronics and Software: An ESP32-based microcontroller board (SparkFun ESP32 thing) was purchased online from SparkFun. The output of the audio amplifier MAX9814 was connected to the analog-to-digital converter of ESP32. The audio signal was sampled at 8 kHz and samples were transmitted to a remote PC over Wi-Fi. In addition to remote transmission, an additional PCB containing an SD card module (Figure S3, Supporting Information) was also designed using the EAGLE design software for off-line recording of digitized samples. The board was ordered from Elecrow Bazaar and surface mount components from Digikey. ECG samples were recorded using AD8232—single lead heart rate monitor—also from Sparkfun. Schematic of the entire circuit is shown in Figure S20 of the Supporting Information. MATLAB was used for creating a graphical interface and data processing. ESP32 was programmed using the PlatformIO integrated development environment with Arduino libraries.

Acquisition of Electrocardiogram: SparkFun's single-lead heart rate monitor (AD8232) sensor and microcontroller (ESP32) were used for the recording of the ECG signals. Simultaneous to ECG recording, the heart sounds using the water-silicone composite transducer were also recorded. The peaks of the ECG signal were detected using a modified version (see Description S2 of the Supporting Information for more information concerning the modifications and comparison of ECG and PCC recordings) of the algorithm described by Pan and Tompkins,^[40] Springer et al.,^[41] and Tosanguan.^[42]

Algorithmic Detection of Heart Rate by PCC: The waveforms recorded with the water-silicone composite transducer between -1 and 1 were normalized and bandpass (passband: 20–150 Hz) filtering to eliminate any background noise was applied. To localize S1 and S2 waveforms, wavelet denoising, envelope detection, and subsequent peak detection algorithms were applied to find the peaks and the distance between S1 signals was calculated to estimate the heart rate (for more information, please see Description S3, Supporting Information).

Supporting Information

Supporting Information is available from the Wiley Online Library or from the author.

Acknowledgements

F.G. would like to thank Imperial College Department of Bioengineering and the Institute for Security Science and Technology (funding under Champions Fund) and the Wellcome Trust (207687/Z/17/Z) for financial support. Y.C. thanks the Turkish Ministry of Education. M.K. acknowledges Finnish Foundation for Technology Promotion. M.K. acknowledges EPSRC DTP (Reference: 1846144). F.G. acknowledges Imperial College Centre for Plastic Electronics and EPSRC for Plastic Electronics Doctoral Training Centre (EP/G037515/1 and EP/L016702/1). F.G. also acknowledges Agri Futures Lab. The authors would like to thank Merve Cirisoglu Cotur for her help with the design of illustrations and videos. The authors would also like to thank Kamila Karimjee for the fruitful discussions and her assistance with the experiments. The Acknowledgements and supplementary videos were updated on April 20, 2020 after initial online publication.

Conflict of Interest

The authors declare no conflict of interest.

Keywords

heart and respiration monitoring, stretchable and flexible materials, wearable devices, wearables for humans and animals, wireless acoustic sensors

Received: December 10, 2019

Revised: January 30, 2020

Published online: February 25, 2020

-
- [1] J. Andreu-Perez, D. R. Leff, H. M. D. Ip, G. Z. Yang, *IEEE Trans. Biomed. Eng.* **2015**, 62, 2750.
- [2] A. Pantelopoulos, N. G. Bourbakis, *IEEE Trans. Syst., Man, Cybern., Part C: Appl. Rev.* **2010**, 40, 1.
- [3] A. Kailas, C. C. Chong, F. Watanabe, *IEEE Pulse* **2010**, 1, 57.
- [4] A. M. Hussain, M. M. Hussain, *Adv. Mater.* **2016**, 28, 4219.
- [5] F. Güder, A. Ainla, J. Redston, B. Mosadegh, A. Glavan, T. J. Martin, G. M. Whitesides, *Angew. Chem., Int. Ed.* **2016**, 55, 5727.
- [6] C. Dincer, R. Bruch, E. Costa-Rama, M. T. Fernández-Abedul, A. Merkoçi, A. Manz, G. A. Urban, F. Güder, *Adv. Mater.* **2019**, 31, 1806739.
- [7] D. Maier, E. Laubender, A. Basavanna, S. Schumann, F. Güder, G. A. Urban, C. Dincer, *ACS Sens.* **2019**, 4, 2945.
- [8] C. Majidi, *Soft Rob.* **2014**, 1, 5.
- [9] X. Zhao, Q. Hua, R. Yu, Y. Zhang, C. Pan, *Adv. Electron. Mater.* **2015**, 1, 1500142.
- [10] S. Sharma, A. El-Laboudi, M. Reddy, N. Jugnee, S. Sivasubramaniyam, S. M. El, P. Georgiou, D. Johnston, N. Oliver, A. E. G. G. Cass, *Anal. Methods* **2018**, 10, 2088.
- [11] A. J. Bandodkar, W. Jia, C. Yardimci, X. Wang, J. Ramirez, J. Wang, *Anal. Chem.* **2015**, 87, 394.
- [12] J. R. Windmiller, J. Wang, *Electroanalysis* **2013**, 25, 29.
- [13] J. H. So, A. S. Tayi, F. Güder, G. M. Whitesides, *Adv. Funct. Mater.* **2014**, 24, 7197.
- [14] O. Ranunkel, F. Güder, H. Arora, *ACS Appl. Bio Mater.* **2019**, 2, 1490.
- [15] X. Wang, Z. Liu, T. Zhang, *Small* **2017**, 13, 1602790.
- [16] E. Jo, K. Lewis, D. Directo, M. J. Y. Kim, B. A. Dolezal, *J. Sport. Sci. Med.* **2016**, 15, 540.
- [17] Z. Zhang, Z. Pi, B. Liu, *IEEE Trans. Biomed. Eng.* **2015**, 62, 522.
- [18] V. Marozas, A. Petrenas, S. Daukantas, A. Lukosevicius, *J. Electrocardiol.* **2011**, 44, 189.
- [19] R. Castrillón, J. J. Pérez, H. Andrade-Caicedo, *Biomed. Eng. Online* **2018**, 17, 38.
- [20] O. T. Inan, P. F. Migeotte, K. S. Park, M. Etemadi, K. Tavakolian, R. Casanella, J. Zanetti, J. Tank, I. Funtova, G. K. Prisk, M. Di Rienzo, *IEEE J. Biomed. Health Inf.* **2015**, 19, 1414.
- [21] S. Yue, H. He, H. Wang, H. Rahul, D. Katabi, *Proc. ACM Interact., Mobile, Wearable Ubiquitous Technol.* **2018**, 2, 22.
- [22] M. Amjadi, K. U. Kyung, I. Park, M. Sitti, *Adv. Funct. Mater.* **2016**, 26, 1678.
- [23] S. Xu, Y. Zhang, L. Jia, K. E. Mathewson, K. I. Jang, J. Kim, H. Fu, X. Huang, P. Chava, R. Wang, S. Bhole, L. Wang, Y. J. Na, Y. Guan, M. Flavin, Z. Han, Y. Huang, J. A. Rogers, *Science* **2014**, 344, 70.
- [24] Y. Liu, J. J. S. Norton, R. Qazi, Z. Zou, K. R. Ammann, H. Liu, L. Yan, P. L. Tran, K. I. Jang, J. W. Lee, D. Zhang, K. A. Kilian, S. H. Jung, T. Bretl, J. Xiao, M. J. Slepian, Y. Huang, J. W. Jeong, J. A. Rogers, *Sci. Adv.* **2016**, 2, e1600167.
- [25] Y. Shu, C. Li, Z. Wang, W. Mi, Y. Li, T. L. Ren, *Sensors* **2015**, 15, 3224.
- [26] P. Chetlur Adithya, R. Sankar, W. A. Moreno, S. Hart, *Biomed. Signal Process. Control* **2017**, 33, 289.
- [27] Y. Nivitha Varghees, K. I. Ramachandran, *IEEE Sens. J.* **2017**, 17, 3861.
- [28] B. S. Emmanuel, *J. Med. Eng. Technol.* **2012**, 36, 303.
- [29] N. Mohammadi-Koushki, H. Memarzadeh-Tehran, S. Goliaei, *41st Conf. on Local Computer Networks (LCN) Workshops*, IEEE Computer Society, Dubai **2016**, pp. 230–235.
- [30] M. Nabih-Ali, E. S. A. El-Dahshan, A. S. Yahia, *J. Med. Eng. Technol.* **2017**, 41, 553.
- [31] Normal Heart Sound, ThinkLabs Medical Library, USA **2016**.
- [32] H. Sakoe, S. Chiba, *IEEE Trans. Acoust., Speech, Signal Process.* **1978**, 26, 43.
- [33] M. Vlachos, M. Hadjieleftheriou, D. Gunopulos, E. Keogh, *VLDB J.* **2006**, 15, 1.
- [34] G. E. A. P. A. Batista, E. J. Keogh, O. M. Taw, V. M. A. De Souza, *Data Min. Knowl. Discovery* **2014**, 28, 634.
- [35] P. Dickens, J. Smith, J. Wolfe, *J. Acoust. Soc. Am.* **2007**, 121, 1471.
- [36] R. A. Casarotto, J. C. Adamowski, F. Fallopa, F. Bacanelli, *Arch. Phys. Med. Rehabil.* **2004**, 85, 162.
- [37] O. A. Godin, *J. Acoust. Soc. Am.* **2008**, 123, 1866.
- [38] K. Vörös, I. Nolte, S. Hungerbühler, J. Reiczigel, J. Ehlers, G. Tater, R. Mischke, T. Zimmering, M. Schneider, *Acta Vet. Hung.* **2011**, 59, 23.
- [39] M. Kasimatis, E. Nunez-Bajo, M. Grell, Y. Cotur, G. Barandun, J.-S. Kim, F. Guder, *ACS Appl. Mater. Interfaces* **2019**, 11, 47577.
- [40] J. Pan, W. J. Tompkins, *IEEE Trans. Biomed. Eng.* **1985**, 32, 230.
- [41] D. B. Springer, T. Brennan, N. Ntusi, H. Y. Abdelrahman, L. J. Zühlke, B. M. Mayosi, L. Tarassenko, G. D. Clifford, *J. Med. Eng. Technol.* **2016**, 40, 342.
- [42] T. Tosanguan, *Development of a Sub-Miniature Acoustic Sensor for Wireless Monitoring of Heart Rate*, Imperial College London, London **2009**.